

10 May 1971
~~11~~

File copy of RFP No. PIC-04-71 (50065) dated
23 Dec 70

Attachment to RFP indicating solicitation of 25X1
dated 23 Dec 70

Copy of the d.o. forwarded with the solicitation
is attached

Copy of letter declining opportunity to
propose, dated 28 Jan 71, is attached. 25X1

Response to telephoned request
by on
10 May 1971. 25X1

25X1

SCHEDULE

DEVELOPMENT OBJECTIVES

OPTICAL EQUIPMENT TEST KIT

1. INTRODUCTION

These development objectives describe the requirements to be met in the design and development of a prototype Optical Equipment Test Kit. Portability, general usefulness, and ease of component replacement are primary considerations.

2. GENERAL DESCRIPTION

This Optical Equipment Test Kit will include those devices commonly and conveniently used during field testing of equipments developed for the purpose of viewing photographic imagery. Such equipments include: microstereoscopes, stereomicroscopes, light tables, projection viewers, and tube magnifiers. Presently, the technical officer monitoring a development contract must collect a miscellaneous assortment of devices such as resolution targets, reticles, and scales in order to field test the performance of an instrument. The purpose of the Optical Equipment Test Kit is to provide a readily portable collection of testing devices for field use during the monitoring of development contracts. It is not the intent that these devices will be capable of the precision of laboratory measurements. In many cases, such as certain optical aberrations, it is sufficient to indicate qualitatively that an anomaly exists. In other cases, such as optical resolution, a more quantitative readout is possible. How quantitative the devices are to be is a trade off between convenience, availability, accuracy, and cost. In all cases, the devices may be either readily available as purchaseable items, or may be specially developed and produced.

3. REQUIREMENTS

3.1 Optical Measurements

The Optical Equipment Test Kit will include devices for determining such optical quantities as field of view, magnification, resolving power (up to 2000 lines per millimeter), astigmatism, field curvature, pin cushion/barrel distortion, and image rotation run-out. It is envisioned that these devices will be a collection of targets, grids, and reticles, either on film or glass substrates. It is expected that some of

these devices will be viewed in the object space of the optics being tested and other devices will be viewed in the eyepiece image plane of the optics being tested. Eyepiece devices should fit a standard Bausch & Lomb 10X wide-field eyepiece, and the eyepiece itself should be calibrated with respect to magnification. Also, some simple form of auxiliary telescope should be included that can be used over the eyepieces of optical systems so as to ensure that resolution readings are limited by the capabilities of the optical system and not limited by the observer's eye.

3.2 Physical Measurements

The Optical Equipment Test Kit will include devices for determining such physical quantities as length, height, temperature, and vibration. These devices would include common English and metric machinists scales, a measuring tape, small internal and external calipers, and a simple depth gauge. Some type of contact thermometer to measure surface temperatures up to 300°F should be included. Also, some type of reticle or target should be included that would indicate the magnitude of mechanical vibration. The intent of this requirement is to view the reticle or target with magnifications and be able to determine the magnitude of the vibrations of the surface on which the reticle or target rests. In addition, grids should be included that would allow checking the accuracy of linear translations and allow checking of orthogonality of an X-Y translations system.

3.3 Miscellaneous Measurements

The Optical Equipment Test Kit will include miscellaneous devices to aid the checking of development equipment. Neutral density filters of 1.0, 1.5, and 2.0 density should be included. A density step wedge of 0.05 to 3.05 density with 21 steps of 0.15 increments is required. In addition, a controllable light source of minimum dimensions is required that can be placed under optics and serve as a source of illumination. Also, a stop watch, camel hair lens and negative brushes, a small flashlight or penlight, some means of measuring web tension of film when spooled on light tables, and a 3-inch or 4-inch hand-held magnifier should also be included.

3.4 Manual

If necessary, a manual should be included in the package to describe the intended use of the devices, give some indication of the probably accuracy of measurement, and outline procedures required for the proper use of the devices.

4. COST REQUIREMENTS

A cost estimate will be submitted for the production of quantities of 5, 10, and 20 units of the Optical Equipment Test Kit.

5. PROPOSALS

Proposals submitted must conform to the attached proposal format.

PROPOSAL FORMAT

All Proposals Must Include the Following Information:

- I. Task Abstract: Contents - Synopsis of task within 12 lines, plus estimated cost of direct labor, material, overhead, G&A, fee, total.
- II. Introduction: Contents - Covering background and task justification rationale.
- III. Technical Discussion: Contents - Detail and subsections as a function of the task.
- IV. Work Statement: Contents - This statement should succinctly describe the individual tasks to be done and should be sufficiently definitive that one may read this section to understand the purpose and scope of the tasks.
- V. Deliverable Items: Contents -
 - 1) Interim and final reports
 - 2) Equipment

25X1

January 28, 1971

CONFIDENTIAL

Day RED

25X1

25X1

Subject: RFP #PIC-04-71(50065)

Gentlemen:

Thank you for affording [] the opportunity to submit a proposal in response to your solicitation. Unfortunately, we are unable to submit a proposal for the development of an optical equipment test kit.

25X1

As [] is a regular manufacturer of a number of potential components for such a test kit, we would be pleased to work with the eventual contractor on this program. The facilities and capabilities of the [] Standards Lab and Precision Ruling Facility are unequalled anywhere. A number of our products might be directly applicable to your requirements. Specifically, special resolution targets, precision grids and scales, diopter telescopes and many other test and inspection aids are available.

25X1

25X1

We regret having to decline the opportunity to propose on this program, however, request that our name be retained on your prospective bidders' list.

Very truly yours,

25X1

[]:sab

[]
Administration Manager
Photogrammetric and Military
Systems

25X1

Group 1
Excluded from automatic
downgrading and
declassification

CONFIDENTIAL